

Gauge Block Set, Metric, Ins. Cert., ISO

Item number: 516-376-10

Images



Mitutoyo Certificate number: 70823309
Date of inspection: 18th Nov. 2008

CERTIFICATE OF INSPECTION

Instrument: Gauge Block
Grade: G (G2M)

Block No.	Material	Material Treatment	Size (mm)	Max. Dev.	Min. Dev.	Surf. Dev.	
1	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
2	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
3	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
4	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
5	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
6	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
7	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
8	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
9	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
10	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
11	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
12	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
13	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
14	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6
15	CrNi12	CrNi12	1.0	-0.4	+1.2	-0.4	1.6

DATE: 18th Nov. 2008



Features

Inch-Metric:	Metric
Type of Certificate:	Inspection certificate
Delivered:	In a wooden case with certificate of inspection
Accuracy:	EN ISO 3650
Grade:	2

Material:

Ceramic

Set configuration:

0,991-0,999 mm (0,001 mm step / 9
blocks)
1,001-1,009 mm (0,001 mm step / 9
blocks)

Mass:

600 g