

# Steel Individual Inch Gauge Block 0.123"

Item number: 614163

## Images



Mitutoyo Certificate number: TMS020000 page 1 of 1

### CERTIFICATE OF INSPECTION

Instrument: Gauge Block  
Grade: G (G10)

Date of inspection: 18th May, 2006 Unit: in, inch

Block	Ident. No.	Central Deviation	Maximum Deviation	Minimum Deviation	Variation
1	06280	-0.4	+1.2	-0.4	1.6

ENVIRONMENT: Air temperature: 68 ± 1.0°F (20 ± 1.0°C)  
 BASE OF REST: 6061-T6 ALUMINUM  
 MATERIAL: Steel  
 INSPECTION METHOD: by comparison  
 COEFFICIENT OF THERMAL EXPANSION: 61.8 x 10<sup>-6</sup> in/in-°F (110.8 x 10<sup>-6</sup> in/in-°C)  
 EXTENDED UNCERTAINTY U<sub>95</sub> for Size: 1.6 in inch (k=2) (Normal length) 1:inch  
 For Gauge Dimension: up to 0.125" (0.2 x 1 / 1.25) in inch  
 REF: 9010  
 QUALITY SYSTEM: ISO 9001 Registered Firm - GB, MICHIGAN  
 TRACEABILITY: Reported in NIST No. 800-20000-02  
 (2000) (National Institute of Standards and Technology)

Date: 18th May, 2006

*S. S. S. S.*  
S. S. S. S.

F-01101  
 0.123 in  
 0.123 in  
 0.123 in

