

## Steel Individual Inch Gauge Block 0.126"

Item number: 614166

### Images



Mitutoyo Certificate number: TMS020000 page 1 of 1

### CERTIFICATE OF INSPECTION

Instrument: Gauge Block  
Grade: G (G10)

Date of inspection: 18th May, 2006 Unit: in, inch

Block	Ident. No.	Central Deviation	Maximum Deviation	Minimum Deviation	Variation
1	06280	-0.4	+1.2	-0.4	1.6

ENVIRONMENT: Air temperature: 68 ± 1.0F (20 ± 1.0 °C)  
 BASE OF REST: 6061 ALUMINUM  
 MATERIAL: Steel  
 INSPECTION METHOD: by comparator  
 COEFFICIENT OF THERMAL EXPANSION: 61.8 x 10<sup>-6</sup> / °F (111.8 x 10<sup>-6</sup> / °C) (10.8 x 10<sup>-6</sup> / °C) (19.4 x 10<sup>-6</sup> / °C)  
 EXTENDED UNCERTAINTY U<sub>95</sub> for Size: 0.4 in, inch (k=2) (k=2)  
 For Gauge Dimension: up to 0.125" (0.2 x 1 / 1.2) in, inch (k=2)  
 QUALITY SYSTEM: ISO 9001 Registered Firm - ISO 9000  
 TRACEABILITY: Traceable to NIST No. 601/2000A-02 (2000/2000A Institute of Standards and Technology)

Date: 18th May, 2006

*S. S. S. S.*

